

Overview of measurement uncertainties for waste oil

The specified measurement uncertainties were in accordance with DIN ISO 11352 (2013-03) determined and correspond to the combined, relative expanded measurement uncertainties ($k=2$, $P=95\%$).

Section 4.2

Parameter	Measuring principle	Norm	Measuring principle
Polychlorinated biphenyls (PCB), per parameter	GC-MS or GC-ECD	DIN EN 12766-1 (2000-11), DIN EN 12766-2 (2001-12) Verfahren B	40 %
Total halogen content	Calorimetric digestion, IC	Anlage 2, Nr. 3 AltölV	15 %